



California Medical Device Recall Information



Recall Name

**Merit Medical Systems Recalls Custom Procedural Trays/Kits
Containing 1% Lidocaine HCl Injection, USP, 10 mg/mL
Due to Oxidized Stainless Steel Particulate**

Recall Date	Product Description	Recalling Firm	Recall Reason
09/30/13	Custom Procedural Trays or Kits <u>containing:</u> 1% Lidocaine HCl Injection, USP, 10 mg/mL Lot # 25-090-DK [Manufactured by Hospira, Inc. (Lake Forest, IL)]	Merit Medical Systems, Inc. Chester, VA	<i>The recalled Custom Procedural Trays or Kits contain a lot of 1% Lidocaine HCl Injection that has been recalled due to the presence of particulates identified as oxidized stainless steel.</i> <i>The presence of particulate in injections can result in life threatening complications.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Affected Merit Medical Lot Numbers for Custom Procedural Trays or Kits	Nationwide	Manufactured from April 29, 2013 through August 20, 2013 Distributed from April 30, 2013 through September 20, 2013

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm389080.htm>